
DURAN® Micro-Burette with Schellbach stripe and glass key, class AS

with straight standard ground stopcock, 30 seconds waiting time, with batch certificate and certificate of conformity





ISO 385 | A 121 °C

With Schellbach stripe and main graduations as circular divisions. Calibration is based on the poured out volume („Ex") at a +20 °C reference temperature. Volume content tolerances conform to DIN.


Typical application: titrations.

Cat. No.	Capacity (mL)	h (mm)	Accuracy limits (mL)	Graduation (mL)	Pack Unit
243201108	1	475	0.01	0.01	1
243201605	2	550	0.01	0.01	1
243202207	5	700	0.01	0.02	1

DURAN® Micro-Burette with Schellbach stripe and PTFE key, class AS

with straight standard ground stopcock, 30 seconds waiting time, with batch certificate and certificate of conformity





ISO 385 | A 121 °C

With Schellbach stripe and main graduations as circular divisions. Calibration is based on the poured out volume („Ex") at a +20 °C reference temperature. Volume content tolerances conform to DIN.

Typical application: titrations.

Cat. No.	Capacity (mL)	l (mm)	Accuracy limits (mL)	Graduation (mL)	Pack Unit
243211109	1	475	0.01	0.01	2
243211606	2	550	0.01	0.01	2
243212208	5	700	0.01	0.02	2
243212705	10	781	0.02	0.02	2